SD Card real time data recorder, Patent CO2, CO, O2, Humidity, Temp. , 6 in 1

AIR QUALITY METER

Model : AQ-9901SD









O2 probe

CO probe





CO₂ probe Humidity probe



NIC The Art of Measurement

SD card real time Data Recorder CO2, CO, O2, Humidity, Dew point, Temp. 6 in 1

AIR QUALITY METER

Model : AQ-9901SD

FEATURES

Ы	LATURES						
*	Real time recorder, save the data into the SD memory						
	card and can be down load to the Excel, extra software						
	is no need. User can make the further data or graphic						
	analysis by themselves. under the Excel software.						
*	At the same time, the SD memory card can record 3						
	probe's data (%RH/CO2/O2/Temp. or						
	%RH/CO2/CO/Temp.) along with the time information						
	into the one Excel file at the same time.						
*	Manual datalogger is available, during execute the						
	manual datalogger function, it can set the different						
	location no. (position 1 to position 99).						
*	Air quality measurement application, multi-function :						
	CO2 (Carbon dioxide), CO (Carbon monoxide), O2						
	(Oxygen in air), Humidity, temperature measurement.						
*	CO2 range : 0 to 4,000 ppm x 1 ppm.						
*	O2 range : 0 to 30.0 % x 0.1 %.						
*	CO range : 0 to 1,000 ppm x 1 ppm.						
*	Humidity range: 10 to 95 %RH.						
*	Dew point Temp. and Wet bulb Temp. measurement.						
*	Temp. range : 0 to 50.0 °C, °C/°F.						
*	CO2 sensor : NDIR, long term reliability.						
	CO, O2 sensor : Galvanic cell type.						
*	Humidity sensor : Precision capacitance sensor						
*	Alarm setting with the beeper sound output.						
*	Sampling time for data recorder is 2 seconds to 8 hours.						
*	Complete set with 4 probes :						
	CO2/Temp. probe, O2/Temp. probe, CO/Temp. probe,						
	Humidity/Temp. probe, main meter and the hard carrying						
	case.						
	Separate probe, easy for remote measurement.						
*	Meter can cooperate with 2 GB to 16 GB SD card, SD						
	card is optional.						
*	RS232/USB computer interface.						
*	Patented.						

Dimension	Meter	177 x 68 x 45 mm		
	Humidity	197 mm in length.		
	probe	, °		
	CO2 probe	190 x 38 x 28 mm		
	O2 probe	150 x 38 x 38 mm		
	CO probe	150 x 38 x 38 mm		
Accessories	Instruction	manual 1 PC		
Included	Hard carryi	ng case, CA-081 PC		
	CO2 probe 1 PC			
	Humidity pr	obe 1 PC		
	O2 probe	1 PC		
	CO probe	1 PC		
Optional	SD memory card (2 GB)			
Accessories	AC to DC 9V adapter.			
	USB cable, USB-01.			
	RS232 cabl	RS232 cable, UPCB-02.		
	Data Acquis	sition software, SW-U801-WIN.		

ELECTRICAL SPECIFICATIONS (23±5°C)

CO2 (Carbon dioxide)

	Range	0 to 4,000 ppm
CO2	Resolution	1 ppm
(Carbon	Accuracy	± 40 ppm
dioxide)	-	* <i>≦</i> 1,000 ppm.
		± 5% of reading
23 ± 5 ℃.		* > 1,000 ppm ≤ 3,000 ppm.
		± 250 ppm typically
		* > 3,000 ppm, reference only
	Repeatability	± 20 ppm * ≤ 3,000 ppm.
Temperature	Range	0 °C to 50 °C,32 °F to 122 °F.
	Resolution	0.1 degree
	Accuracy	°C : ± 0.8 °C °F : ± 1.5 °F.

GENERAL SPECIFICATIONS

1

GENERAL SPEC					
Circuit		one-chip of microprocessor LSI			
	circuit.				
Display		: 52 mm x 38 mm			
		green backlight (ON/OFF).			
Measurement	CO2 (Car	bon dioxide)			
		bon monoxide)			
		gen in air)			
	Humidity				
		nt Temp., Wet bulb Temp.			
	Tempera				
Sensor		NDIR * Nondispersive infrared sensor			
structure		Precision capacitance sensor			
structure		Galvanic cell type			
	CO	Galvanic cell type			
Detelement	Temp.	Precision thermistor			
Datalogger	Auto	2 sec to 8 hour 59 min. 59 sec.			
Sampling Time		@ Sampling time can set to 1 second,			
Setting range		but memory data may loss.			
	Manual	Push the data logger button			
		once will save data one time.			
		@ Set the sampling time to			
		0 second.			
		@ Manual mode, can also select the			
		1 to 99 position (Location) no.			
Data error no.	0.1% of	total saved data max.			
Memory Card		pry card. 1 GB to 16 GB.			
Advanced		nory card Format			
setting	* Set cloc				
-	* Set sam	pling time			
@ main setting	* Auto po	wer OFF management			
		p Sound ON/OFF			
	* Decima	point of SD card setting			
	* Temp. u	unit setting			
	* Alarm v	* Alarm value setting			
Data Hold	Freeze th	ne display reading.			
Memory Recall	Maximun	n & Minimum value.			
Sampling Time	Approx.	1 second.			
of Display					
Data Output	RS 232/L	JSB PC computer interface.			
		ct the optional RS232 cable			
		UPCB-02 will get the RS232 plug.			
		* Connect the optional USB cable			
		1 will get the USB plug.			
Power Supply	* Alkalin	e or heavy duty DC 1.5 V battery			
Tower Supply	(LIM2	AA) x 6 PCs or oquivalant			
	* DC OV	AA) x 6 PCs, or equivalent. adapter input. (AC/DC power			
D 0 1		er is optional).			
Power Current	CO2	Normal operation (w/o SD card save			
	measure-	data and LCD Backlight is OFF) :			
	ment	Approx. DC 136.5 mA.			
		When SD card save the data and LCD			
		Backlight is OFF) :			
		Approx. DC 166 mA.			
	Humidity	Normal operation (w/o SD card save			
	measure-	data and LCD Backlight is OFF) :			
	ment	Approx. DC 10.5 mA.			
	ment				
		When SD card save the data and LCD			
		Backlight is OFF) :			
	-	Approx. DC 40 mA.			
	O2 or	Normal operation (w/o SD card save			
	со	data and LCD Backlight is OFF) :			
	measure-	Approx. DC 12.5 mA.			
	ment	When SD card save the data and LCD			
		Backlight is OFF) :			
		Approx. DC 42.5 mA.			
	* If I CD	backlight on, the power			
		nption will increase approx.			
Onentin	12 mA				
Operating	U TO 50 (C. (32 to 122 °F).			
Temperature		848/ B 11			
	Il ess that	1 80% R.H.			
Operating					
Operating Humidity					
	350 g/0.	77 LB.			

CO (Carbon dioxide)

	Range	0 to 1,000 ppm		
00	Resolution	1 ppm		
* Carbon	Accuracy	± (5% + 2 ppm)		
monoxide	Response	< 30 seconds		
	time *			
	Repeatability	< 2%		
	Zero drift in	< 5 ppm		
	long term			
	Sensitivity	< 5% per year		
	drift			
	* The response time value is specified to reach the 90% reading value.			
Temperature	Range	0 °C to 50 °C ,32 °F to 122 °F.		
	Resolution	0.1 degree		
	Accuracy	°C : ± 0.8 °C °F : ± 1.5 °F.		

O2 (Air oxygen)

	Range	0 to 30 %O2.
02	Resolution	0.1 %02.
* Air oxygen	Accuracy	± (1% reading + 0.2% 02).
	-	@ After calibration
	Response time	≦ 15 seconds. @ t 90
	Overload	100 %02.
	protection	
		0.9 to 1.1 atmosphere.
	pressure range	
	Expected life	≥2 years.
	time	
		If the measurement Air oxygen
		value < 18.0 %O2, the buzzer
		will sound for warning.
emperature	Range	0 °C to 50 °C,32 °F to 122 °F.
	Resolution	0.1 degree
	Accuracy	$^{\circ}C : \pm 0.8 \ ^{\circ}C$ $^{\circ}F : \pm 1.5 \ ^{\circ}F.$

Humidity/Temperature

	Range	5 % to	5 % to 95 % R.H.		
Humidity	Resolution	0.1 %	R.H.		
-	Accuracy	≥70%	≧70% RH :		
		± (3	± (3% reading + 1% RH).		
		< 70% RH :			
		±	3% RH.		
	Range	0 °C to	50 °C,32 °F to 122 °F.		
Temperature	Resolution	0.1 degree			
	Accuracy	°C	± 0.8 ℃.		
		°F	± 1.5 °F.		

Dew Point Temp. (Humidity)

°C	Range	-25.3 °C to 48.9 °C		
	Resolution	0.1 ℃		
°F	Range	-13.5 °F to 120.1 °F.		
	Resolution	0.1 °F.		
Remark :				
* Dew Point display value is calculated from the				
Humidity/Temp. measurement automatically.				
* The Dew Point accuracy is sum accuracy value of Humidity				
& Tempera	ature measureme	nt		
Wet bulb T	emp. (Humid	ity)		
1				

	°C	Range	-21.6 °C to 50.0 °C		
		Resolution	0.1 °C		
	°F Range		-6.9 °F to 122.0 °F.		
		Resolution	0.1 °F.		
Remark :					
	* Wet bulb display value is calculated from the Humidity/Temp.				
	measurement automatically.				
	* The Welt bulb accuracy is sum accuracy value of Humidity				
1	& Temperature measurement				
t to change without notice. 1			110629-A0	29901SD	
	5				